

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102422\
 Data File : VX032351.D
 Acq On : 24 Oct 2022 19:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 03:59:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	47.715	4.6	98	0.00
3 P	Chloromethane	50.000	44.896	10.2	85	0.00
4 C	Vinyl Chloride	50.000	45.594	8.8#	88	0.00
5 T	Bromomethane	50.000	53.584	-7.2	103	0.00
6 T	Chloroethane	50.000	36.900	26.2#	76	0.00
7 T	Trichlorofluoromethane	50.000	52.166	-4.3	105	0.00
8 T	Diethyl Ether	50.000	51.473	-2.9	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.443	1.1	101	0.00
10 T	Methyl Iodide	50.000	35.482	29.0#	67	0.00
11 T	Tert butyl alcohol	250.000	256.204	-2.5	101	0.00
12 CM	1,1-Dichloroethene	50.000	48.857	2.3#	94	0.00
13 T	Acrolein	250.000	288.691	-15.5	115	0.00
14 T	Allyl chloride	50.000	51.882	-3.8	99	0.00
15 T	Acrylonitrile	250.000	253.594	-1.4	97	0.00
16 T	Acetone	250.000	255.467	-2.2	102	0.00
17 T	Carbon Disulfide	50.000	46.875	6.3	89	0.00
18 T	Methyl Acetate	50.000	52.553	-5.1	100	0.00
19 T	Methyl tert-butyl Ether	50.000	53.937	-7.9	104	0.00
20 T	Methylene Chloride	50.000	46.204	7.6	94	0.00
21 T	trans-1,2-Dichloroethene	50.000	48.794	2.4	93	0.00
22 T	Diisopropyl ether	50.000	52.972	-5.9	103	0.00
23 T	Vinyl Acetate	250.000	263.531	-5.4	99	0.00
24 P	1,1-Dichloroethane	50.000	51.256	-2.5	100	0.00
25 T	2-Butanone	250.000	261.944	-4.8	101	0.00
26 T	2,2-Dichloropropane	50.000	54.655	-9.3	101	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.080	-4.2	99	0.00
28 T	Bromochloromethane	50.000	47.564	4.9	99	0.00
29 T	Tetrahydrofuran	250.000	254.819	-1.9	97	0.00
30 C	Chloroform	50.000	53.242	-6.5#	103	0.00
31 T	Cyclohexane	50.000	48.911	2.2	97	0.00
32 T	1,1,1-Trichloroethane	50.000	56.739	-13.5	107	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.934	-5.9	102	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	102	0.00
35 S	Dibromofluoromethane	50.000	51.273	-2.5	100	0.00
36 T	1,1-Dichloropropene	50.000	49.858	0.3	98	0.00
37 T	Ethyl Acetate	50.000	49.914	0.2	99	0.00
38 T	Carbon Tetrachloride	50.000	54.769	-9.5	105	0.00
39 T	Methylcyclohexane	50.000	47.936	4.1	96	0.00
40 TM	Benzene	50.000	49.000	2.0	96	0.00
41 T	Methacrylonitrile	50.000	53.959	-7.9	99	0.00
42 TM	1,2-Dichloroethane	50.000	53.725	-7.5	104	0.00
43 T	Isopropyl Acetate	50.000	53.762	-7.5	101	0.00
44 TM	Trichloroethene	50.000	49.360	1.3	96	0.00
45 C	1,2-Dichloropropane	50.000	51.236	-2.5#	99	0.00
46 T	Dibromomethane	50.000	52.410	-4.8	101	0.00
47 T	Bromodichloromethane	50.000	57.246	-14.5	109	0.00
48 T	Methyl methacrylate	50.000	54.439	-8.9	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032351.D
 Acq On : 24 Oct 2022 19:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 25 03:59:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	985.391	1.5	92	0.00
50 S	Toluene-d8	50.000	49.149	1.7	93	0.00
51 T	4-Methyl-2-Pentanone	250.000	269.931	-8.0	101	0.00
52 CM	Toluene	50.000	49.902	0.2#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	51.600	-3.2	106	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.796	-9.6	102	0.00
55 T	1,1,2-Trichloroethane	50.000	52.930	-5.9	102	0.00
56 T	Ethyl methacrylate	50.000	55.099	-10.2	104	0.00
57 T	1,3-Dichloropropane	50.000	52.529	-5.1	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	236.718	5.3	92	0.00
59 T	2-Hexanone	250.000	267.712	-7.1	101	0.00
60 T	Dibromochloromethane	50.000	59.910	-19.8	110	0.00
61 T	1,2-Dibromoethane	50.000	53.655	-7.3	102	0.00
62 S	4-Bromofluorobenzene	50.000	53.761	-7.5	101	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	102	0.00
64 T	Tetrachloroethene	50.000	48.786	2.4	97	0.00
65 PM	Chlorobenzene	50.000	50.909	-1.8	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.582	-11.2	104	0.00
67 C	Ethyl Benzene	50.000	52.196	-4.4#	100	0.00
68 T	m/p-Xylenes	100.000	103.042	-3.0	100	0.00
69 T	o-Xylene	50.000	52.980	-6.0	101	0.00
70 T	Styrene	50.000	53.409	-6.8	102	0.00
71 P	Bromoform	50.000	54.992	-10.0	114	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	105	0.00
73 T	Isopropylbenzene	50.000	50.438	-0.9	102	0.00
74 T	N-amyl acetate	50.000	53.345	-6.7	102	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.409	-0.8	102	0.00
76 T	1,2,3-Trichloropropane	50.000	52.553	-5.1	106	0.00
77 T	Bromobenzene	50.000	50.402	-0.8	103	0.00
78 T	n-propylbenzene	50.000	51.328	-2.7	102	0.00
79 T	2-Chlorotoluene	50.000	50.710	-1.4	103	0.00
80 T	1,3,5-Trimethylbenzene	50.000	50.660	-1.3	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	50.940	-1.9	106	0.00
82 T	4-Chlorotoluene	50.000	51.504	-3.0	103	0.00
83 T	tert-Butylbenzene	50.000	51.234	-2.5	104	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.783	-5.6	104	0.00
85 T	sec-Butylbenzene	50.000	51.878	-3.8	104	0.00
86 T	p-Isopropyltoluene	50.000	52.188	-4.4	103	0.00
87 T	1,3-Dichlorobenzene	50.000	49.702	0.6	102	0.00
88 T	1,4-Dichlorobenzene	50.000	49.707	0.6	102	0.00
89 T	n-Butylbenzene	50.000	51.018	-2.0	100	0.00
90 T	Hexachloroethane	50.000	56.304	-12.6	111	0.00
91 T	1,2-Dichlorobenzene	50.000	51.527	-3.1	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	58.574	-17.1	109	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.936	2.1	98	0.00
94 T	Hexachlorobutadiene	50.000	46.598	6.8	96	0.00
95 T	Naphthalene	50.000	51.789	-3.6	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
Data File : VX032351.D
Acq On : 24 Oct 2022 19:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 25 03:59:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	49.217	1.6	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6