

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032351.D
 Acq On : 24 Oct 2022 19:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/25/2022
 Supervised By :Mahesh Dadoda 10/26/2022

Quant Time: Oct 25 03:59:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125658	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198463	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	180027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86531	52.934	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 105.860%			
35) Dibromofluoromethane	5.391	113	70810	51.273	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.540%			
50) Toluene-d8	8.653	98	244335	49.149	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 98.300%			
62) 4-Bromofluorobenzene	11.079	95	97868	53.761	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.520%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	54853	47.715	ug/l	97
3) Chloromethane	1.295	50	50194	44.896	ug/l	97
4) Vinyl Chloride	1.374	62	56452	45.594	ug/l	99
5) Bromomethane	1.612	94	69464	53.584	ug/l	97
6) Chloroethane	1.685	64	41089	36.900	ug/l	95
7) Trichlorofluoromethane	1.886	101	114820	52.166	ug/l	100
8) Diethyl Ether	2.136	74	38501	51.473	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	59000	49.443	ug/l	98
10) Methyl Iodide	2.453	142	61187	35.482	ug/l	97
11) Tert butyl alcohol	2.977	59	58850	256.204	ug/l	97
12) 1,1-Dichloroethene	2.319	96	54798	48.857	ug/l	93
13) Acrolein	2.239	56	57448	288.691	ug/l	98
14) Allyl chloride	2.666	41	91393	51.882	ug/l	99
15) Acrylonitrile	3.063	53	155474	253.594	ug/l	99
16) Acetone	2.386	43	135522	255.467	ug/l	99
17) Carbon Disulfide	2.514	76	123082	46.875	ug/l	100
18) Methyl Acetate	2.703	43	94648	52.553	ug/l	98
19) Methyl tert-butyl Ether	3.111	73	208459	53.937	ug/l	99
20) Methylene Chloride	2.788	84	62936	46.204	ug/l	99
21) trans-1,2-Dichloroethene	3.093	96	59607	48.794	ug/l	98
22) Diisopropyl ether	3.764	45	204340	52.972	ug/l	91
23) Vinyl Acetate	3.721	43	812505	263.531	ug/l	100
24) 1,1-Dichloroethane	3.611	63	114240	51.256	ug/l	99
25) 2-Butanone	4.556	43	216952	261.944	ug/l	98
26) 2,2-Dichloropropane	4.477	77	87766	54.655	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	74490	52.080	ug/l	99
28) Bromochloromethane	4.904	49	43737	47.564	ug/l	99
29) Tetrahydrofuran	5.007	42	136311	254.819	ug/l	99
30) Chloroform	5.099	83	129159	53.242	ug/l	100
31) Cyclohexane	5.477	56	91403	48.911	ug/l	99
32) 1,1,1-Trichloroethane	5.385	97	117377	56.739	ug/l	99
36) 1,1-Dichloropropene	5.696	75	88572	49.858	ug/l	99
37) Ethyl Acetate	4.715	43	86820	49.914	ug/l	100
38) Carbon Tetrachloride	5.678	117	102713	54.769	ug/l	99
39) Methylcyclohexane	7.385	83	101731	47.936	ug/l	97
40) Benzene	6.044	78	252296	49.000	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032351.D
 Acq On : 24 Oct 2022 19:26
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 10/25/2022
 Supervised By :Mahesh Dadoda 10/26/2022

Quant Time: Oct 25 03:59:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	49491	53.959	ug/l	97
42) 1,2-Dichloroethane	6.086	62	105957	53.725	ug/l	99
43) Isopropyl Acetate	6.342	43	146771	53.762	ug/l	100
44) Trichloroethene	7.129	130	69986	49.360	ug/l	98
45) 1,2-Dichloropropane	7.434	63	66550	51.236	ug/l	99
46) Dibromomethane	7.580	93	50773	52.410	ug/l	98
47) Bromodichloromethane	7.824	83	102446	57.246	ug/l	96
48) Methyl methacrylate	7.696	41	74119	54.439	ug/l	98
49) 1,4-Dioxane	7.684	88	26925	985.391	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	461935	269.931	ug/l	100
52) Toluene	8.720	92	166442	49.902	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	104178	51.600	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	109268	54.796	ug/l	99
55) 1,1,2-Trichloroethane	9.153	97	71717	52.930	ug/l	97
56) Ethyl methacrylate	9.116	69	107521	55.099	ug/l	98
57) 1,3-Dichloropropane	9.311	76	115286	52.529	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.245	63	211733	236.718	ug/l	100
59) 2-Hexanone	9.433	43	341845	267.712	ug/l	100
60) Dibromochloromethane	9.525	129	83195	59.910	ug/l	99
61) 1,2-Dibromoethane	9.610	107	77068	53.655	ug/l	98
64) Tetrachloroethene	9.275	164	57996	48.786	ug/l	97
65) Chlorobenzene	10.080	112	187905	50.909	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.165	131	73408	55.582	ug/l	99
67) Ethyl Benzene	10.195	91	341091	52.196	ug/l	100
68) m/p-Xylenes	10.305	106	265923	103.042	ug/l	98
69) o-Xylene	10.640	106	130983	52.980	ug/l	100
70) Styrene	10.659	104	222667	53.409	ug/l	99
71) Bromoform	10.799	173	61868	54.992	ug/l #	100
73) Isopropylbenzene	10.964	105	352022	50.438	ug/l	100
74) N-amyl acetate	10.842	43	137547	53.345	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	113109	50.409	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	101851m	52.553	ug/l	
77) Bromobenzene	11.201	156	86225	50.402	ug/l	98
78) n-propylbenzene	11.305	91	410199	51.328	ug/l	100
79) 2-Chlorotoluene	11.366	91	244243	50.710	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	305400	50.660	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	31454	50.940	ug/l	98
82) 4-Chlorotoluene	11.457	91	288812	51.504	ug/l	100
83) tert-Butylbenzene	11.720	119	309028	51.234	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	309362	52.783	ug/l	99
85) sec-Butylbenzene	11.890	105	378160	51.878	ug/l	99
86) p-Isopropyltoluene	12.012	119	323692	52.188	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	163653	49.702	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	168127	49.707	ug/l	99
89) n-Butylbenzene	12.335	91	278711	51.018	ug/l	99
90) Hexachloroethane	12.543	117	54189	56.304	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	166558	51.527	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	25110	58.574	ug/l	96
93) 1,2,4-Trichlorobenzene	13.591	180	102976	48.936	ug/l	99
94) Hexachlorobutadiene	13.725	225	42558	46.598	ug/l	99
95) Naphthalene	13.774	128	349046	51.789	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	104171	49.217	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
Data File : VX032351.D
Acq On : 24 Oct 2022 19:26
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022

Quant Time: Oct 25 03:59:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 733
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:168 Resp: 125658
Ion Ratio Lower Upper
168 100
99 59.0 44.8 67.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#2
Dichlorodifluoromethane
Concen: 47.715 ug/l
RT: 1.167 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 85 Resp: 54853
Ion Ratio Lower Upper
85 100
87 32.1 16.8 50.4





#3
Chloromethane
Concen: 44.896 ug/l
RT: 1.295 min Scan# 34
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 50 Resp: 50194
Ion Ratio Lower Upper
50 100
52 32.7 24.7 37.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#4
Vinyl Chloride
Concen: 45.594 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 62 Resp: 56452
Ion Ratio Lower Upper
62 100
64 33.8 26.6 39.8





#5
Bromomethane
Concen: 53.584 ug/l
RT: 1.612 min Scan# 80
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 69464
Ion Ratio Lower Upper
94 100
96 93.3 76.7 115.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#6
Chloroethane
Concen: 36.900 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 64 Resp: 41089
Ion Ratio Lower Upper
64 100
66 30.6 26.6 39.8





#7
Trichlorofluoromethane
Concen: 52.166 ug/l
RT: 1.886 min Scan# 114820
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:101 Resp: 114820
Ion Ratio Lower Upper
101 100
103 64.3 51.3 76.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#8
Diethyl Ether
Concen: 51.473 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 74 Resp: 38501
Ion Ratio Lower Upper
74 100
45 93.5 45.9 137.6





#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.443 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:101 Resp: 59000
Ion Ratio Lower Upper
101 100
85 46.8 36.5 54.7
151 80.7 65.5 98.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#10
Methyl Iodide
Concen: 35.482 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion:142 Resp: 61187
Ion Ratio Lower Upper
142 100
127 49.0 36.9 55.3
141 14.6 11.4 17.2





#11

Tert butyl alcohol

Concen: 256.204 ug/l

RT: 2.977 min Scan# 310

Delta R.T. 0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 59 Resp: 58850

Ion Ratio Lower Upper

59 100





57 10.4 7.5 11.3

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#12

1,1-Dichloroethene

Concen: 48.857 ug/l

RT: 2.319 min Scan# 202

Delta R.T. -0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26




Tgt Ion: 96 Resp: 54798


Ion Ratio Lower Upper

96 100

61 167.0 124.0 186.0


98 62.7 50.0 75.0





#13
Acrolein
Concen: 288.691 ug/l
RT: 2.239 min Scan# 188
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 57448
Ion Ratio Lower Upper
56 100
55 72.3 56.4 84.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#14
Allyl chloride
Concen: 51.882 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 41 Resp: 91393
Ion Ratio Lower Upper
41 100
39 75.8 59.5 89.3
76 35.5 28.6 42.8





#15
Acrylonitrile
Concen: 253.594 ug/l
RT: 3.063 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 155474
Ion Ratio Lower Upper
53 100
52 83.2 65.9 98.9
51 37.4 29.7 44.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#16
Acetone
Concen: 255.467 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 43 Resp: 135522
Ion Ratio Lower Upper
43 100
58 29.6 24.2 36.2









#17
Carbon Disulfide
Concen: 46.875 ug/l
RT: 2.514 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 76 Resp: 12308
Ion Ratio Lower Upper
76 100
78 9.6 7.6 11.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#18
Methyl Acetate
Concen: 52.553 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 43 Resp: 94648
Ion Ratio Lower Upper
43 100
74 24.0 19.8 29.8









#19
Methyl tert-butyl Ether
Concen: 53.937 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 73 Resp: 20845
Ion Ratio Lower Upper
73 100
57 21.4 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#20
Methylene Chloride
Concen: 46.204 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 84 Resp: 62936
Ion Ratio Lower Upper
84 100
49 116.3 92.6 139.0
51 36.7 29.2 43.8
86 63.2 51.4 77.0





#21
trans-1,2-Dichloroethene
Concen: 48.794 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 96 Resp: 5960
Ion Ratio Lower Upper
96 100
61 131.8 107.7 161.5
98 62.1 50.2 75.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#22
Diisopropyl ether
Concen: 52.972 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 45 Resp: 204340
Ion Ratio Lower Upper
45 100
43 72.4 67.5 101.3
87 31.0 24.0 36.0
59 12.0 9.7 14.5









#23
 Vinyl Acetate
 Concen: 263.531 ug/l
 RT: 3.721 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX032351.D
 Acq: 24 Oct 2022 19:26

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion: 43 Resp: 81250
 Ion Ratio Lower Upper
 43 100
 86 10.7 8.6 13.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 10/25/2022
 Supervised By :Mahesh Dadoda 10/26/2022



#24
 1,1-Dichloroethane
 Concen: 51.256 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.000 min
 Lab File: VX032351.D
 Acq: 24 Oct 2022 19:26

Tgt Ion: 63 Resp: 114240
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.6 10.9
 100 4.6 2.1 6.3









#25
2-Butanone
Concen: 261.944 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 43 Resp: 216952
Ion Ratio Lower Upper
43 100
72 28.0 21.7 32.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#26
2,2-Dichloropropane
Concen: 54.655 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 77 Resp: 87766
Ion Ratio Lower Upper
77 100
97 22.8 11.8 35.3





#27

cis-1,2-Dichloroethene

Concen: 52.080 ug/l

RT: 4.495 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 96 Resp: 74490





Ion	Ratio	Lower	Upper
96	100		
61	141.4	0.0	279.2
98	64.7	0.0	128.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#28

Bromochloromethane

Concen: 47.564 ug/l

RT: 4.904 min Scan# 626




Delta R.T. 0.006 min


Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Tgt Ion: 49 Resp: 43737

Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	4.8
130	84.8	68.3	102.5





#29

Tetrahydrofuran

Concen: 254.819 ug/l

RT: 5.007 min Scan# 643

Delta R.T. 0.000 min

Lab File: VX032351.D


Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion:	42	Resp:	13631
Ion	Ratio	Lower	Upper
42	100		
72	47.9	37.5	56.3
71	44.0	34.7	52.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#30

Chloroform

Concen: 53.242 ug/l




RT: 5.099 min Scan# 658


Delta R.T. 0.006 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26


Tgt Ion:	83	Resp:	129159
Ion	Ratio	Lower	Upper
83	100		
85	65.8	52.5	78.7





#31
Cyclohexane
Concen: 48.911 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 56 Resp: 9140
Ion Ratio Lower Upper
56 100
69 32.1 25.4 38.0
84 90.1 71.2 106.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#32
1,1,1-Trichloroethane
Concen: 56.739 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 97 Resp: 117377
Ion Ratio Lower Upper
97 100
99 63.9 51.8 77.8
61 44.3 35.7 53.5





#33

1,2-Dichloroethane-d4

Concen: 52.934 ug/l

RT: 5.958 min Scan# 798

Delta R.T. 0.006 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 65 Resp: 8653

Ion Ratio Lower Upper

65 100





67 50.2 0.0 103.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26




Tgt Ion:114 Resp: 198463


Ion Ratio Lower Upper

114 100

63 20.5 0.0 38.6


88 14.9 0.0 32.6





#35
Dibromofluoromethane
Concen: 51.273 ug/l
RT: 5.391 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:113 Resp: 70810
Ion Ratio Lower Upper
113 100
111 100.7 82.8 124.2
192 18.7 15.5 23.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#36
1,1-Dichloropropene
Concen: 49.858 ug/l
RT: 5.696 min Scan# 756
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion: 75 Resp: 88572
Ion Ratio Lower Upper
75 100
110 36.2 18.4 55.0
77 31.5 24.7 37.1





#37

Ethyl Acetate

Concen: 49.914 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 86820





Ion	Ratio	Lower	Upper
43	100		
61	14.2	11.4	17.2
70	10.9	8.8	13.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#38

Carbon Tetrachloride

Concen: 54.769 ug/l

RT: 5.678 min Scan# 753




Delta R.T. -0.000 min


Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Tgt Ion:117 Resp: 102713

Ion	Ratio	Lower	Upper
117	100		
119	95.2	75.4	113.0
121	30.3	23.9	35.9





#39

Methylcyclohexane

Concen: 47.936 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 10173

Ion Ratio Lower Upper

83 100

55 74.3 56.6 85.0





98 47.5 37.4 56.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#40

Benzene

Concen: 49.000 ug/l

RT: 6.044 min Scan# 813

Delta R.T. 0.006 min

Lab File: VX032351.D




Acq: 24 Oct 2022 19:26


Tgt Ion: 78 Resp: 252296

Ion Ratio Lower Upper

78 100

77 23.4 18.6 27.8






#41
Methacrylonitrile
Concen: 53.959 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC






Tgt Ion: 41 Resp: 4949
Ion Ratio Lower Upper
41 100
39 59.3 45.5 68.3
67 71.7 59.4 89.2
52 32.3 25.3 37.9

Manual Integrations

APPROVED




Reviewed By :Krupa Patel 10/25/2022


Supervised By :Mahesh Dadoda 10/26/2022



#42
1,2-Dichloroethane
Concen: 53.725 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 62 Resp: 105957
Ion Ratio Lower Upper
62 100
98 8.4 0.0 17.4





#43
Isopropyl Acetate
Concen: 53.762 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 14677
Ion Ratio Lower Upper
43 100
61 19.2 15.5 23.3
87 14.0 11.0 16.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#44
Trichloroethene
Concen: 49.360 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion:130 Resp: 69986
Ion Ratio Lower Upper
130 100
95 100.2 0.0 196.8









#45
1,2-Dichloropropane
Concen: 51.236 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 63 Resp: 66550
Ion Ratio Lower Upper
63 100
65 29.5 23.8 35.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#46
Dibromomethane
Concen: 52.410 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 93 Resp: 50773
Ion Ratio Lower Upper
93 100
95 81.3 66.1 99.1
174 98.0 80.7 121.1





#47
Bromodichloromethane
Concen: 57.246 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 102440
Ion Ratio Lower Upper
83 100
85 63.1 53.0 79.6
127 9.0 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#48
Methyl methacrylate
Concen: 54.439 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 41 Resp: 74119
Ion Ratio Lower Upper
41 100
69 82.5 67.3 100.9
39 66.2 53.8 80.8





#49

1,4-Dioxane

Concen: 985.391 ug/l

RT: 7.684 min Scan# 1082

Delta R.T. -0.006 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 26925





Ion	Ratio	Lower	Upper
88	100		
43	32.2	26.9	40.3
58	65.1	50.7	76.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#50

Toluene-d8

Concen: 49.149 ug/l

RT: 8.653 min Scan# 1241




Delta R.T. -0.000 min


Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Tgt Ion: 98 Resp: 244335


Ion	Ratio	Lower	Upper
98	100		
100	64.3	50.7	76.1





#51
4-Methyl-2-Pentanone
Concen: 269.931 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 461938
Ion Ratio Lower Upper
43 100
58 38.5 31.0 46.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#52
Toluene
Concen: 49.902 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion: 92 Resp: 166442
Ion Ratio Lower Upper
92 100
91 176.6 140.3 210.5





#53

t-1,3-Dichloropropene

Concen: 51.600 ug/l

RT: 8.982 min Scan# 1129

Delta R.T. 0.006 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 75 Resp: 104178

Ion Ratio Lower Upper

75 100





77 31.6 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#54

cis-1,3-Dichloropropene

Concen: 54.796 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26




Tgt Ion: 75 Resp: 109268


Ion Ratio Lower Upper

75 100

77 32.3 25.3 37.9


39 53.9 42.4 63.6





#55
1,1,2-Trichloroethane
Concen: 52.930 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion	Ratio	Lower	Upper
97	100		
83	83.8	68.8	103.2
85	54.5	44.7	67.1
99	60.0	50.6	76.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#56
Ethyl methacrylate
Concen: 55.099 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion	Ratio	Lower	Upper
69	100		
41	67.0	54.5	81.7
39	43.3	35.7	53.5





#57

1,3-Dichloropropane

Concen: 52.529 ug/l

RT: 9.311 min Scan# 11528

Delta R.T. 0.006 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 11528

Ion Ratio Lower Upper

76 100





78 32.4 26.0 39.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#58

2-Chloroethyl Vinyl ether

Concen: 236.718 ug/l

RT: 8.245 min Scan# 1174

Delta R.T. -0.000 min

Lab File: VX032351.D




Acq: 24 Oct 2022 19:26


Tgt Ion: 63 Resp: 211733

Ion Ratio Lower Upper

63 100


106 29.4 23.6 35.4





#59
2-Hexanone
Concen: 267.712 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 34184
Ion Ratio Lower Upper
43 100
58 53.4 26.7 80.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#60
Dibromochloromethane
Concen: 59.910 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion:129 Resp: 83195
Ion Ratio Lower Upper
129 100
127 76.2 38.4 115.2









#61
1,2-Dibromoethane
Concen: 53.655 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:107 Resp: 77068
Ion Ratio Lower Upper
107 100
109 93.5 76.5 114.7




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#62
4-Bromofluorobenzene
Concen: 53.761 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion: 95 Resp: 97868
Ion Ratio Lower Upper
95 100
174 80.3 0.0 166.8
176 78.3 0.0 161.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:117 Resp: 18002
Ion Ratio Lower Upper
117 100
82 56.1 44.2 66.4
119 32.3 25.8 38.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#64
Tetrachloroethene
Concen: 48.786 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion:164 Resp: 57996
Ion Ratio Lower Upper
164 100
166 125.9 104.9 157.3
129 95.8 77.8 116.6
131 93.9 76.3 114.5









#65
Chlorobenzene
Concen: 50.909 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:112 Resp: 18790
Ion Ratio Lower Upper
112 100
114 31.0 25.7 38.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 55.582 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion:131 Resp: 73408
Ion Ratio Lower Upper
131 100
133 96.3 47.6 142.8
119 64.5 31.7 95.1





#67
Ethyl Benzene
Concen: 52.196 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 34109
Ion Ratio Lower Upper
91 100
106 31.1 25.0 37.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#68
m/p-Xylenes
Concen: 103.042 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.006 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion:106 Resp: 265923
Ion Ratio Lower Upper
106 100
91 205.9 162.6 244.0









#69
o-Xylene
Concen: 52.980 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 130983
Ion Ratio Lower Upper
106 100
91 216.9 108.3 324.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#70
Styrene
Concen: 53.409 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion:104 Resp: 222667
Ion Ratio Lower Upper
104 100
78 52.1 40.8 61.2
103 55.0 44.2 66.2





#71

Bromoform

Concen: 54.992 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:173 Resp: 6186

Ion Ratio Lower Upper

173 100

175 48.3 24.1 72.4





254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26




Tgt Ion:152 Resp: 97812


Ion Ratio Lower Upper

152 100

115 86.1 43.1 129.4


150 173.0 0.0 348.6





#73
Isopropylbenzene
Concen: 50.438 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion:105 Resp: 35202
Ion Ratio Lower Upper
105 100
120 26.3 13.2 39.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#74
N-ethyl acetate
Concen: 53.345 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 43 Resp: 137547
Ion Ratio Lower Upper
43 100
70 42.0 33.9 50.9
55 24.3 19.1 28.7
61 23.3 19.0 28.4









#75
1,1,2,2-Tetrachloroethane
Concen: 50.409 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.001 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 113109
Ion Ratio Lower Upper
83 100
131 10.5 5.2 15.6
85 65.0 32.3 96.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#76
1,2,3-Trichloropropane
Concen: 52.553 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 75 Resp: 101851
Ion Ratio Lower Upper
75 100
77 36.6 16.0 47.9





#77

Bromobenzene

Concen: 50.402 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC





Tgt Ion:	156	Resp:	86225
Ion Ratio	Lower	Upper	
156	100		
77	155.6	75.5	226.5
158	97.4	49.0	147.0

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#78

n-propylbenzene

Concen: 51.328 ug/l




RT: 11.305 min Scan# 1676


Delta R.T. -0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Tgt Ion:	91	Resp:	410199
Ion Ratio	Lower	Upper	
91	100		
120	23.4	11.6	34.8









#79
2-Chlorotoluene
Concen: 50.710 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 91 Resp: 24424
Ion Ratio Lower Upper
91 100
126 34.3 17.3 51.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#80
1,3,5-Trimethylbenzene
Concen: 50.660 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion:105 Resp: 305400
Ion Ratio Lower Upper
105 100
120 48.2 23.5 70.6





#81
trans-1,4-Dichloro-2-butene
Concen: 50.940 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 31454
Ion Ratio Lower Upper
75 100
53 111.2 87.8 131.8
89 49.9 37.8 56.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#82
4-Chlorotoluene
Concen: 51.504 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26

Tgt Ion: 91 Resp: 288812
Ion Ratio Lower Upper
91 100
126 29.7 14.9 44.5





#83

tert-Butylbenzene

Concen: 51.234 ug/l

RT: 11.720 min Scan# 1744

Delta R.T. 0.007 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:119 Resp: 309028





Ion	Ratio	Lower	Upper
119	100		
91	58.5	29.1	87.3
134	22.2	11.3	33.8

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#84

1,2,4-Trimethylbenzene

Concen: 52.783 ug/l

RT: 11.750 min Scan# 1749




Delta R.T. -0.000 min


Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Tgt Ion:105 Resp: 309362

Ion	Ratio	Lower	Upper
105	100		
120	43.2	22.1	66.1









#85
 sec-Butylbenzene
 Concen: 51.878 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX032351.D
 Acq: 24 Oct 2022 19:26

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tgt Ion:105 Resp: 378160
 Ion Ratio Lower Upper
 105 100
 134 19.7 10.1 30.3




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 10/25/2022
 Supervised By :Mahesh Dadoda 10/26/2022



#86
 p-Isopropyltoluene
 Concen: 52.188 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX032351.D
 Acq: 24 Oct 2022 19:26


Tgt Ion:119 Resp: 323692
 Ion Ratio Lower Upper
 119 100
 134 25.6 13.2 39.5
 91 24.9 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 49.702 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 16365
Ion Ratio Lower Upper
146 100
111 41.0 20.3 61.0
148 64.7 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#88
1,4-Dichlorobenzene
Concen: 49.707 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Tgt Ion:146 Resp: 168127
Ion Ratio Lower Upper
146 100
111 40.2 19.8 59.4
148 64.8 31.9 95.9





#89
n-Butylbenzene
Concen: 51.018 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 27871
Ion Ratio Lower Upper
91 100
92 51.7 26.3 78.9
134 25.9 13.0 39.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#90
Hexachloroethane
Concen: 56.304 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26


Tgt Ion:117 Resp: 54189
Ion Ratio Lower Upper
117 100
201 75.9 38.5 115.3





#91
1,2-Dichlorobenzene
Concen: 51.527 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 166558
Ion Ratio Lower Upper
146 100
111 41.9 21.1 63.1
148 63.9 32.3 96.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022




#92
1,2-Dibromo-3-Chloropropane
Concen: 58.574 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion: 75 Resp: 25110
Ion Ratio Lower Upper
75 100
155 83.3 43.1 129.3
157 109.1 57.1 171.5





#93

1,2,4-Trichlorobenzene

Concen: 48.936 ug/l

RT: 13.591 min Scan# 2051

Delta R.T. 0.000 min

Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC

Tgt Ion:180 Resp: 102970





Ion	Ratio	Lower	Upper
180	100		
182	94.7	46.8	140.3
145	31.8	15.6	46.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 10/25/2022

Supervised By :Mahesh Dadoda 10/26/2022



#94

Hexachlorobutadiene

Concen: 46.598 ug/l

RT: 13.725 min Scan# 2073




Delta R.T. 0.000 min


Lab File: VX032351.D

Acq: 24 Oct 2022 19:26

Tgt Ion:225 Resp: 42558


Ion	Ratio	Lower	Upper
225	100		
223	63.8	31.4	94.3
227	63.7	32.0	96.0





#95
Naphthalene
Concen: 51.789 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 349040
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 10.9 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 10/25/2022
Supervised By :Mahesh Dadoda 10/26/2022



#96
1,2,3-Trichlorobenzene
Concen: 49.217 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX032351.D
Acq: 24 Oct 2022 19:26



Tgt Ion:180 Resp: 104171
Ion Ratio Lower Upper
180 100
182 95.4 47.4 142.2
145 33.9 16.5 49.5

