



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102422\
 Data File : VX032327.D
 Acq On : 24 Oct 2022 09:46
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Title : SW846 8260
 Last Update : Sat Oct 22 04:56:40 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	6974	PASS
75	95	30	60	52.9	18799	PASS
95	95	100	100	100.0	35547	PASS
96	95	5	9	6.5	2328	PASS
173	174	0.00	2	0.7	198	PASS
174	95	50	100	82.5	29312	PASS
175	174	5	9	7.3	2134	PASS
176	174	95	101	98.0	28725	PASS
177	176	5	9	6.7	1930	PASS