

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
 Data File : VX038450.D
 Acq On : 24 Oct 2023 12:39
 Operator : JC/MD
 Sample : VX1024WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1024WBL01

Quant Time: Oct 25 06:07:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142565	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	253236	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251148	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129757	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	97031	42.315	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.620%
35) Dibromofluoromethane	5.385	113	82293	45.099	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.200%
50) Toluene-d8	8.647	98	309607	44.403	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.800%#
62) 4-Bromofluorobenzene	11.079	95	117317	44.381	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
Data File : VX038450.D
Acq On : 24 Oct 2023 12:39
Operator : JC/MD
Sample : VX1024WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1024WBL01

Quant Time: Oct 25 06:07:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

