

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
 Data File : VX038453.D
 Acq On : 24 Oct 2023 13:54
 Operator : JC/MD
 Sample : 04936-07ME
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1084ME

Quant Time: Oct 24 15:00:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	157061	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	272024	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	269998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	100354	39.725	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	79.440%		
35) Dibromofluoromethane	5.391	113	79772	40.698	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	81.400%		
50) Toluene-d8	8.647	98	329355	43.972	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	87.940%#		
62) 4-Bromofluorobenzene	11.085	95	131205	46.206	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	92.420%		
Target Compounds						
16) Acetone	2.422	43	25148	23.852	ug/l	# 87
52) Toluene	8.720	92	28597	5.807	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
Data File : VX038453.D
Acq On : 24 Oct 2023 13:54
Operator : JC/MD
Sample : 04936-07ME
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1084ME

Quant Time: Oct 24 15:00:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

