

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
 Data File : VX038454.D
 Acq On : 24 Oct 2023 14:45
 Operator : JC/MD
 Sample : 04975-03 10X
 Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 61

Quant Time: Oct 25 06:08:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	150307	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.757	114	266911	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	270411	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99972	41.352	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.700%
35) Dibromofluoromethane	5.385	113	83204	43.262	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.520%
50) Toluene-d8	8.647	98	325660	44.312	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.620%#
62) 4-Bromofluorobenzene	11.079	95	129603	46.516	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.040%
Target Compounds						
					Qvalue	
31) Cyclohexane	5.544	56	5254	2.006	ug/l #	43
52) Toluene	8.720	92	38417	7.950	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
Data File : VX038454.D
Acq On : 24 Oct 2023 14:45
Operator : JC/MD
Sample : 04975-03 10X
Misc : 1.09g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
61

Quant Time: Oct 25 06:08:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

