

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
 Data File : VX038455.D
 Acq On : 24 Oct 2023 15:08
 Operator : JC/MD
 Sample : 04975-02 10X
 Misc : 1.26g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 56

Manual Integrations
 APPROVED

Reviewed By :John Carlone 10/25/2023
 Supervised By :Mahesh Dadoda 10/25/2023

Quant Time: Oct 25 06:08:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	161198	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	283547	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288794	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103026	39.736	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	79.480%		
35) Dibromofluoromethane	5.385	113	91089	44.583	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.160%		
50) Toluene-d8	8.647	98	349665	44.787	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	89.580%#		
62) 4-Bromofluorobenzene	11.079	95	142712	48.216	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.440%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	4915	4.542	ug/l	94
31) Cyclohexane	5.501	56	9289m	3.307	ug/l	
39) Methylcyclohexane	7.385	83	14179	4.296	ug/l	# 87
52) Toluene	8.720	92	350257	68.233	ug/l	95
67) Ethyl Benzene	10.195	91	49894	4.625	ug/l	98
68) m/p-Xylenes	10.299	106	63924	15.326	ug/l	98
69) o-Xylene	10.640	106	31278	7.683	ug/l	99
73) Isopropylbenzene	10.963	105	19067	1.754	ug/l	99
78) n-propylbenzene	11.305	91	58667	4.415	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	48057	5.125	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	162066	17.336	ug/l	97
85) sec-Butylbenzene	11.890	105	24254	2.044	ug/l	87
86) p-Isopropyltoluene	12.006	119	11148	1.138	ug/l	94
89) n-Butylbenzene	12.329	91	34609	3.549	ug/l	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
Data File : VX038455.D
Acq On : 24 Oct 2023 15:08
Operator : JC/MD
Sample : 04975-02 10X
Misc : 1.26g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
56

Quant Time: Oct 25 06:08:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/25/2023
Supervised By : Mahesh Dadoda 10/25/2023

