

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
 Data File : VX038457.D
 Acq On : 24 Oct 2023 15:54
 Operator : JC/MD
 Sample : 04975-04
 Misc : 1.11g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1061

Quant Time: Oct 25 06:09:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154669	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	271878	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	270793	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153955	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99622	40.045	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.080%
35) Dibromofluoromethane	5.391	113	83029	42.382	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	84.760%
50) Toluene-d8	8.647	98	329691	44.041	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.080%#
62) 4-Bromofluorobenzene	11.085	95	129554	45.649	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
Data File : VX038457.D
Acq On : 24 Oct 2023 15:54
Operator : JC/MD
Sample : 04975-04
Misc : 1.11g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1061

Quant Time: Oct 25 06:09:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

