

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
 Data File : VX038458.D
 Acq On : 24 Oct 2023 16:18
 Operator : JC/MD
 Sample : 05022-01
 Misc : 1.13g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OILY-DEBRIS

Quant Time: Oct 25 06:09:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	158567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	274841	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	276786	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158135	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	99688	39.086	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	78.180%
35) Dibromofluoromethane	5.391	113	82169	41.491	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	82.980%
50) Toluene-d8	8.647	98	335016	44.270	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.540%#
62) 4-Bromofluorobenzene	11.085	95	133114	46.398	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102423\
Data File : VX038458.D
Acq On : 24 Oct 2023 16:18
Operator : JC/MD
Sample : 05022-01
Misc : 1.13g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OILY-DEBRIS

Quant Time: Oct 25 06:09:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

