

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102425\
 Data File : VX048349.D
 Acq On : 24 Oct 2025 13:07
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 24 23:04:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 20 14:13:59 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	148473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	292353	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	292509	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	142629	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	116641	56.032	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.060%
35) Dibromofluoromethane	5.373	113	79776	39.970	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	79.940%
50) Toluene-d8	8.635	98	336954	48.165	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.340%
62) 4-Bromofluorobenzene	11.061	95	138430	52.014	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102425\
Data File : VX048349.D
Acq On : 24 Oct 2025 13:07
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 24 23:04:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102025W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 20 14:13:59 2025
Response via : Initial Calibration

