

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102518\
 Data File : VX005584.D
 Acq On : 25 Oct 2018 12:57
 Operator : JC/MD
 Sample : J5655-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-EB-2018-1

Quant Time: Oct 26 05:58:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	205443	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	308254	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	270475	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	130927	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	129194	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
35) Dibromofluoromethane	5.50	113	98975	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
50) Toluene-d8	8.71	98	359369	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
62) 4-Bromofluorobenzene	11.14	95	120824	42.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.32%	
Target Compounds						
16) Acetone	2.44	43	2853	2.153	ug/l	99
20) Methylene Chloride	2.86	84	45947	20.702	ug/l	96
25) 2-Butanone	4.70	43	6897	3.418	ug/l #	90
29) Tetrahydrofuran	5.15	42	162201	123.511	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102518\
Data File : VX005584.D
Acq On : 25 Oct 2018 12:57
Operator : JC/MD
Sample : J5655-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-EB-2018-1

Quant Time: Oct 26 05:58:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

