

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102518\
 Data File : VX005585.D
 Acq On : 25 Oct 2018 13:23
 Operator : JC/MD
 Sample : J5655-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0006

Quant Time: Oct 26 06:02:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	199228	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	302610	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	274102	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	133978	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	129832	53.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.78%	
35) Dibromofluoromethane	5.49	113	103578	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
50) Toluene-d8	8.71	98	369058	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
62) 4-Bromofluorobenzene	11.14	95	122581	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
Target Compounds						
16) Acetone	2.45	43	7563	5.885	ug/l	97
39) Methylcyclohexane	7.46	83	8601	2.652	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102518\
Data File : VX005585.D
Acq On : 25 Oct 2018 13:23
Operator : JC/MD
Sample : J5655-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0006

Quant Time: Oct 26 06:02:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

