

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102518\
 Data File : VX005594.D
 Acq On : 25 Oct 2018 17:23
 Operator : JC/MD
 Sample : J5665-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-03

Quant Time: Oct 26 06:58:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 25 02:52:45 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	207670	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	313658	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	281478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	135432	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	138011	54.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	5.51	113	108643	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
50) Toluene-d8	8.71	98	375519	49.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.42%	
62) 4-Bromofluorobenzene	11.14	95	129145	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
Target Compounds						
16) Acetone	2.45	43	4630	3.456	ug/l #	85
67) Ethyl Benzene	10.25	91	27325	2.771	ug/l	97
68) m/p-Xylenes	10.36	106	76020	19.869	ug/l	99
69) o-Xylene	10.70	106	30235	8.400	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102518\
Data File : VX005594.D
Acq On : 25 Oct 2018 17:23
Operator : JC/MD
Sample : J5665-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-03

Quant Time: Oct 26 06:58:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102418W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 25 02:52:45 2018
Response via : Initial Calibration

