

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013205.D
 Acq On : 24 Oct 2019 12:27
 Operator : JC/SP
 Sample : K5531-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20191021

Quant Time: Oct 25 03:09:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	81732	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	120526	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	118087	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	77239	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	86553	50.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.26%	
35) Dibromofluoromethane	5.48	113	66746	49.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.32%	
50) Toluene-d8	8.71	98	242619	48.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.42%	
62) 4-Bromofluorobenzene	11.13	95	90174	45.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.48%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013205.D
Acq On : 24 Oct 2019 12:27
Operator : JC/SP
Sample : K5531-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20191021

Quant Time: Oct 25 03:09:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

