

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013210.D
 Acq On : 24 Oct 2019 14:42
 Operator : JC/SP
 Sample : K5532-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0673

Quant Time: Oct 25 04:10:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	80648	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	117509	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	115920	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	73769	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	87472	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
35) Dibromofluoromethane	5.49	113	67107	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
50) Toluene-d8	8.71	98	243719	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	11.14	95	88020	45.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.58%	
Target Compounds						
4) Vinyl Chloride	1.40	62	10261	7.007	ug/l	90
12) 1,1-Dichloroethene	2.36	96	709	0.516	ug/l #	78
16) Acetone	2.44	43	1194	1.502	ug/l #	81
17) Carbon Disulfide	2.56	76	2114	0.572	ug/l	96
21) trans-1,2-Dichloroethene	3.15	96	64435	45.124	ug/l	91
27) cis-1,2-Dichloroethene	4.59	96	443630	263.669	ug/l	88
44) Trichloroethene	7.21	130	20653	12.329	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013210.D
Acq On : 24 Oct 2019 14:42
Operator : JC/SP
Sample : K5532-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0673

Quant Time: Oct 25 04:10:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

