

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013211.D
 Acq On : 24 Oct 2019 15:06
 Operator : JC/SP
 Sample : K5532-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0672

Quant Time: Oct 29 12:13:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	79582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	114925	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	114032	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	71645	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	84638	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
35) Dibromofluoromethane	5.49	113	64683	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	8.71	98	233969	49.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.52%	
62) 4-Bromofluorobenzene	11.13	95	84466	44.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.86%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	7315	9.328	ug/l	96
17) Carbon Disulfide	2.57	76	2505	0.687	ug/l	95
21) trans-1,2-Dichloroethene	3.16	96	3635	3.028	ug/l #	80
27) cis-1,2-Dichloroethene	4.59	96	5326	3.208	ug/l	83
44) Trichloroethene	7.22	130	2097	1.280	ug/l	86
52) Toluene	8.78	92	1946	0.509	ug/l	94
95) Naphthalene	13.83	128	2607	0.487	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013211.D
Acq On : 24 Oct 2019 15:06
Operator : JC/SP
Sample : K5532-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0672

Quant Time: Oct 29 12:13:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

