

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013212.D
 Acq On : 24 Oct 2019 15:29
 Operator : JC/SP
 Sample : K5532-01DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0670DL

Quant Time: Oct 25 04:42:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	81019	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	119502	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	116649	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	73563	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	87297	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
35) Dibromofluoromethane	5.49	113	66797	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
50) Toluene-d8	8.71	98	241675	48.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.88%	
62) 4-Bromofluorobenzene	11.14	95	88847	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.40	62	1328	0.903	ug/l #	81
16) Acetone	2.44	43	914	1.145	ug/l #	69
21) trans-1,2-Dichloroethene	3.16	96	18003	12.893	ug/l	90
27) cis-1,2-Dichloroethene	4.59	96	45066	26.662	ug/l	88
44) Trichloroethene	7.21	130	107972	63.380	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013212.D
Acq On : 24 Oct 2019 15:29
Operator : JC/SP
Sample : K5532-01DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0670DL

Quant Time: Oct 25 04:42:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

