

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013213.D
 Acq On : 24 Oct 2019 15:52
 Operator : JC/SP
 Sample : K5532-06DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0673DL

Quant Time: Oct 25 04:46:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	78334	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	115022	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	110025	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	69810	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	90468	55.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.44%	
35) Dibromofluoromethane	5.49	113	68560	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
50) Toluene-d8	8.71	98	251094	52.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.64%	
62) 4-Bromofluorobenzene	11.13	95	89794	47.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.46%	
Target Compounds						
4) Vinyl Chloride	1.40	62	1877	1.320	ug/l #	76
21) trans-1,2-Dichloroethene	3.16	96	12096	9.105	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	83680	51.204	ug/l	88
44) Trichloroethene	7.21	130	3812	2.325	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013213.D
Acq On : 24 Oct 2019 15:52
Operator : JC/SP
Sample : K5532-06DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0673DL

Quant Time: Oct 25 04:46:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

