

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013216.D
 Acq On : 24 Oct 2019 17:02
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Oct 26 02:05:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	78091	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	114580	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	113289	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	71255	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	88583	54.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.48%	
35) Dibromofluoromethane	5.49	113	67986	53.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.42%	
50) Toluene-d8	8.71	98	244246	51.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.16%	
62) 4-Bromofluorobenzene	11.14	95	91974	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.56	76	2029	0.567	ug/l	100
21) trans-1,2-Dichloroethene	3.16	96	2196	2.047	ug/l #	81
27) cis-1,2-Dichloroethene	4.58	96	2275	1.396	ug/l	92
44) Trichloroethene	7.20	130	6916	4.234	ug/l	99
95) Naphthalene	13.84	128	1773	0.333	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013216.D
Acq On : 24 Oct 2019 17:02
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Oct 26 02:05:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

