

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013225.D
 Acq On : 25 Oct 2019 12:58
 Operator : JC/SP
 Sample : K5547-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Oct 25 23:56:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	73467	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	108854	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	107739	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	62912	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	83711	54.48	ug/l	0.00
Spiked Amount			Recovery	=	108.96%	
35) Dibromofluoromethane	5.49	113	63244	52.10	ug/l	0.00
Spiked Amount			Recovery	=	104.20%	
50) Toluene-d8	8.71	98	226705	50.40	ug/l	0.00
Spiked Amount			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.14	95	79401	44.59	ug/l	0.00
Spiked Amount			Recovery	=	89.18%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013225.D
Acq On : 25 Oct 2019 12:58
Operator : JC/SP
Sample : K5547-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Oct 25 23:56:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

