

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013230.D
 Acq On : 25 Oct 2019 14:54
 Operator : JC/SP
 Sample : K5546-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01-102419

Quant Time: Oct 26 00:22:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	70743	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	107671	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	104111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	62310	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	81817	55.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.60%	
35) Dibromofluoromethane	5.49	113	61874	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
50) Toluene-d8	8.71	98	220446	49.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.08%	
62) 4-Bromofluorobenzene	11.13	95	76924	43.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.36%	
Target Compounds						
16) Acetone	2.44	43	5229	7.501	ug/l	90
20) Methylene Chloride	2.84	84	6323	5.111	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013230.D
Acq On : 25 Oct 2019 14:54
Operator : JC/SP
Sample : K5546-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01-102419

Quant Time: Oct 26 00:22:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

