

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013232.D
 Acq On : 25 Oct 2019 15:40
 Operator : JC/SP
 Sample : K5546-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW810-8.0-13.0-102419

Quant Time: Oct 26 00:36:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	78056	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	117822	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	117781	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	70356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	88922	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
35) Dibromofluoromethane	5.49	113	66287	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
50) Toluene-d8	8.71	98	235980	48.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.92%	
62) 4-Bromofluorobenzene	11.14	95	84461	43.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.66%	
Target Compounds						
16) Acetone	2.43	43	5335	6.936	ug/l	98
52) Toluene	8.78	92	1886	0.481	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013232.D
Acq On : 25 Oct 2019 15:40
Operator : JC/SP
Sample : K5546-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW810-8.0-13.0-102419

Quant Time: Oct 26 00:36:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

