

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013235.D
 Acq On : 25 Oct 2019 16:50
 Operator : JC/SP
 Sample : K5532-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0672

Quant Time: Oct 29 12:12:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	70444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	105607	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	102426	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	64943	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	82259	55.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.66%	
35) Dibromofluoromethane	5.49	113	63047	53.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.08%	
50) Toluene-d8	8.71	98	222620	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
62) 4-Bromofluorobenzene	11.14	95	80846	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	6504	9.370	ug/l	98
17) Carbon Disulfide	2.56	76	4344	1.346	ug/l	97
18) Methyl Acetate	2.76	43	1730	2.088	ug/l	91
21) trans-1,2-Dichloroethene	3.16	96	3140	2.967	ug/l	86
27) cis-1,2-Dichloroethene	4.58	96	3953	2.690	ug/l	93
44) Trichloroethene	7.22	130	585	0.389	ug/l	41
52) Toluene	8.78	92	1870	0.532	ug/l	95
95) Naphthalene	13.83	128	4516	0.931	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013235.D
Acq On : 25 Oct 2019 16:50
Operator : JC/SP
Sample : K5532-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0672

Quant Time: Oct 29 12:12:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration

