

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX102519\
 Data File : VX013225.D
 Acq On : 25 Oct 2019 12:58
 Operator : JC/SP
 Sample : K5547-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Oct 25 23:56:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 24 03:56:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	73467	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	108854	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	107739	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	62912	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	83711	54.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.96%	
35) Dibromofluoromethane	5.49	113	63244	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
50) Toluene-d8	8.71	98	226705	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.14	95	79401	44.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.18%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX102519\
Data File : VX013225.D
Acq On : 25 Oct 2019 12:58
Operator : JC/SP
Sample : K5547-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Oct 25 23:56:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X102419W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 24 03:56:38 2019
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.65 min Scan# 781
Delta R.T. 0.00 min
Lab File: VX013225.D
Acq: 25 Oct 2019 12:58

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Tot Ion:168 Resp: 73467
Ion Ratio Lower Upper
168 100
99 64.4 47.0 70.4




Abundance Ion 167.90 (167.60 to 168.60): V
Ion 98.90 (98.60 to 99.60): VX0






#33
1,2-Dichloroethane-d4
Concen: 54.478 ug/l
RT: 6.05 min Scan# 847
Delta R.T. 0.00 min
Lab File: VX013225.D
Acq: 25 Oct 2019 12:58

Tgt Ion: 65 Resp: 83711
Ion Ratio Lower Upper
65 100
67 50.2 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0
Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.85 min Scan# 978
 Delta R.T. 0.00 min
 Lab File: VX013225.D
 Acq: 25 Oct 2019 12:58

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Tot Ion:114 Resp: 108854
 Ion Ratio Lower Upper
 114 100
 63 20.0 0.0 39.0
 88 17.1 0.0 32.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 52.101 ug/l
 RT: 5.49 min Scan# 755
 Delta R.T. 0.01 min
 Lab File: VX013225.D
 Acq: 25 Oct 2019 12:58

Tgt Ion:113 Resp: 63244
 Ion Ratio Lower Upper
 113 100
 111 103.8 80.3 120.5
 192 20.8 15.9 23.9



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.396 ug/l
RT: 8.71 min Scan# 1283
Delta R.T. 0.00 min
Lab File: VX013225.D
Acq: 25 Oct 2019 12:58




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA


Tot Ion: 98 Resp: 226705
Ion Ratio Lower Upper
98 100
100 66.3 53.3 79.9



#62
4-Bromofluorobenzene
Concen: 44.594 ug/l
RT: 11.14 min Scan# 1681
Delta R.T. 0.01 min
Lab File: VX013225.D
Acq: 25 Oct 2019 12:58

Tgt Ion: 95 Resp: 79401
Ion Ratio Lower Upper
95 100
174 80.7 0.0 152.4
176 75.1 0.0 146.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1513
Delta R.T. 0.00 min
Lab File: VX013225.D
Acq: 25 Oct 2019 12:58

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Tot Ion:117 Resp: 107739
Ion Ratio Lower Upper
117 100
82 56.4 46.5 69.7
119 32.0 25.2 37.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.07 min Scan# 1835
Delta R.T. 0.00 min
Lab File: VX013225.D
Acq: 25 Oct 2019 12:58

Tgt Ion:152 Resp: 62912
Ion Ratio Lower Upper
152 100
115 64.8 46.6 139.8
150 158.3 0.0 354.0

