

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
 Data File : VX024920.D
 Acq On : 25 Oct 2021 19:23
 Operator : JC/MD
 Sample : M4304-01
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P056-TW001-01

Quant Time: Oct 26 01:26:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 26 01:24:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	77842	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	71262	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	33822	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	7837	3.429	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.600%
7) Chloroethane-d5	1.673	69	8883	4.314	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.200%
11) 1,1-Dichloroethene-d2	2.313	63	23841	2.799	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.000%#
20) 2-Butanone-d5	4.477	46	40860	48.904	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.800%
24) Chloroform-d	5.068	84	48624	4.420	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.964	65	24859	4.134	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.600%
32) Benzene-d6	5.983	84	75997	4.577	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.600%
36) 1,2-Dichloropropane-d6	7.312	67	22157	4.183	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.600%
41) Toluene-d8	8.653	98	71743	4.310	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	8.958	79	9986	4.107	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	9.385	63	40362	46.124	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.240%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	18687	4.615	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.400%
66) 1,2-Dichlorobenzene-d4	12.323	152	24993	4.385	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	87.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
Data File : VX024920.D
Acq On : 25 Oct 2021 19:23
Operator : JC/MD
Sample : M4304-01
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P056-TW001-01

Quant Time: Oct 26 01:26:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 26 01:24:18 2021
Response via : Initial Calibration

