

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
 Data File : VX024929.D
 Acq On : 25 Oct 2021 22:58
 Operator : JC/MD
 Sample : M4318-07
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P020-TW001-02

Quant Time: Oct 26 01:28:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 26 01:24:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	77786	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	72672	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	32847	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	7806	3.418	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	68.400%
7) Chloroethane-d5	1.672	69	8921	4.336	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.800%
11) 1,1-Dichloroethene-d2	2.313	63	23238	2.730	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.600%#
20) 2-Butanone-d5	4.483	46	39219	46.973	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.940%
24) Chloroform-d	5.062	84	48558	4.417	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.400%
26) 1,2-Dichloroethane-d4	5.964	65	25548	4.252	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.000%
32) Benzene-d6	5.983	84	75172	4.439	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.800%
36) 1,2-Dichloropropane-d6	7.318	67	22782	4.217	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.400%
41) Toluene-d8	8.653	98	71001	4.182	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.600%
43) trans-1,3-Dichloroprop...	8.958	79	9482	3.824	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.400%
46) 2-Hexanone-d5	9.391	63	39924	44.738	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.480%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	18460	4.471	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.400%
66) 1,2-Dichlorobenzene-d4	12.323	152	25646	4.633	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
Data File : VX024929.D
Acq On : 25 Oct 2021 22:58
Operator : JC/MD
Sample : M4318-07
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P020-TW001-02

Quant Time: Oct 26 01:28:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 26 01:24:18 2021
Response via : Initial Calibration

