

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
 Data File : VX024931.D
 Acq On : 25 Oct 2021 23:45
 Operator : JC/MD
 Sample : M4304-03
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P060-TW001-01

Quant Time: Oct 26 01:28:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 26 01:24:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	78012	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	70230	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	33794	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	7530	3.288	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	65.800%
7) Chloroethane-d5	1.672	69	7478	3.624	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	72.400%
11) 1,1-Dichloroethene-d2	2.307	63	21958	2.572	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	51.400%#
20) 2-Butanone-d5	4.483	46	36728	43.862	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.720%
24) Chloroform-d	5.068	84	46070	4.178	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.600%
26) 1,2-Dichloroethane-d4	5.964	65	24656	4.092	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.800%
32) Benzene-d6	5.983	84	71571	4.374	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.400%
36) 1,2-Dichloropropane-d6	7.312	67	21512	4.121	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	82.400%
41) Toluene-d8	8.653	98	67615	4.121	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.400%
43) trans-1,3-Dichloroprop...	8.958	79	9827	4.101	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	9.391	63	38201	44.296	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.600%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	17567	4.403	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	12.323	152	24237	4.256	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
Data File : VX024931.D
Acq On : 25 Oct 2021 23:45
Operator : JC/MD
Sample : M4304-03
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P060-TW001-01

Quant Time: Oct 26 01:28:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 26 01:24:18 2021
Response via : Initial Calibration

