

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102521\  
 Data File : VX024924.D  
 Acq On : 25 Oct 2021 20:58  
 Operator : JC/MD  
 Sample : M4304-02  
 Misc : 25.0mL/MSVOA\_X/WATER  
 ALS Vial : 13 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 P059-TW001-01

Quant Time: Oct 26 01:27:19 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\SFAMXTR102221WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Tue Oct 26 01:24:18 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 6.769  | 114   | 79794    | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 10.055 | 117   | 74990    | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 12.030 | 152   | 35555    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.368  | 65    | 6757     | 2.884    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 57.600%  |
| 7) Chloroethane-d5            | 1.673  | 69    | 7575     | 3.589    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 71.800%  |
| 11) 1,1-Dichloroethene-d2     | 2.307  | 63    | 20857    | 2.388    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 47.800%# |
| 20) 2-Butanone-d5             | 4.477  | 46    | 36273    | 42.352   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 84.700%  |
| 24) Chloroform-d              | 5.068  | 84    | 43288    | 3.838    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 76.800%  |
| 26) 1,2-Dichloroethane-d4     | 5.964  | 65    | 22603    | 3.667    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 73.400%  |
| 32) Benzene-d6                | 5.977  | 84    | 65947    | 3.774    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 75.400%  |
| 36) 1,2-Dichloropropane-d6    | 7.312  | 67    | 20832    | 3.737    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 74.800%  |
| 41) Toluene-d8                | 8.653  | 98    | 63796    | 3.642    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 72.800%  |
| 43) trans-1,3-Dichloroprop... | 8.952  | 79    | 8580     | 3.353    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 67.000%  |
| 46) 2-Hexanone-d5             | 9.391  | 63    | 36299    | 39.419   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 78.840%  |
| 56) 1,1,2,2-Tetrachloroeth... | 11.195 | 84    | 16878    | 3.961    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 79.200%  |
| 66) 1,2-Dichlorobenzene-d4    | 12.323 | 152   | 21792    | 3.637    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 72.800%# |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102521\  
Data File : VX024924.D  
Acq On : 25 Oct 2021 20:58  
Operator : JC/MD  
Sample : M4304-02  
Misc : 25.0mL/MSVOA\_X/WATER  
ALS Vial : 13 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
P059-TW001-01

Quant Time: Oct 26 01:27:19 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\SFAMXTR102221WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Tue Oct 26 01:24:18 2021  
Response via : Initial Calibration

