

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
 Data File : VX024928.D
 Acq On : 25 Oct 2021 22:34
 Operator : JC/MD
 Sample : M4318-06
 Misc : 25.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P017-TW001-02

Quant Time: Oct 26 01:28:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 26 01:24:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	72848	5.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	68569	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	31062	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	7983	3.733	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.600%
7) Chloroethane-d5	1.672	69	8871	4.604	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.000%
11) 1,1-Dichloroethene-d2	2.313	63	23205	2.911	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	58.200%#
20) 2-Butanone-d5	4.477	46	39608	50.655	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.300%
24) Chloroform-d	5.068	84	48740	4.734	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
26) 1,2-Dichloroethane-d4	5.964	65	25145	4.468	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.400%
32) Benzene-d6	5.983	84	73845	4.622	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.400%
36) 1,2-Dichloropropane-d6	7.318	67	22141	4.344	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	8.653	98	69162	4.318	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.400%
43) trans-1,3-Dichloroprop...	8.958	79	8726	3.730	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.600%
46) 2-Hexanone-d5	9.391	63	39056	46.384	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.760%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	18973	4.870	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.400%
66) 1,2-Dichlorobenzene-d4	12.323	152	24761	4.730	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102521\
Data File : VX024928.D
Acq On : 25 Oct 2021 22:34
Operator : JC/MD
Sample : M4318-06
Misc : 25.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P017-TW001-02

Quant Time: Oct 26 01:28:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXTR102221WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 26 01:24:18 2021
Response via : Initial Calibration

