

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102522\  
 Data File : VX032355.D  
 Acq On : 25 Oct 2022 11:28  
 Operator : JC/MD  
 Sample : VX1025WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX1025WBL01

Quant Time: Oct 27 14:16:00 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X101922W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Oct 22 04:56:40 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 134991   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 210389   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 184720   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 86630    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 84148    | 47.918   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 95.840%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 67686    | 46.233   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 92.460%  |
| 50) Toluene-d8              | 8.653  | 98    | 248688   | 47.189   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 94.380%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 85880    | 44.502   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 89.000%  |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX102522\  
Data File : VX032355.D  
Acq On : 25 Oct 2022 11:28  
Operator : JC/MD  
Sample : VX1025WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX1025WBL01

Quant Time: Oct 27 14:16:00 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X101922W.M  
Quant Title : SW846 8260  
QLast Update : Sat Oct 22 04:56:40 2022  
Response via : Initial Calibration

