

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032364.D
 Acq On : 25 Oct 2022 15:13
 Operator : JC/MD
 Sample : N5249-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-32-20221019

Quant Time: Oct 27 14:17:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209292	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	182292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85551	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84094	48.346	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.700%
35) Dibromofluoromethane	5.385	113	67449	46.312	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.620%
50) Toluene-d8	8.653	98	247806	47.268	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.540%
62) 4-Bromofluorobenzene	11.079	95	85203	44.382	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.760%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.325	96	350	0.293	ug/l #	82
16) Acetone	2.386	43	1193	2.113	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	2610	1.715	ug/l	89
44) Trichloroethene	7.129	130	167207	111.826	ug/l	99
52) Toluene	8.726	92	2804	0.797	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032364.D
Acq On : 25 Oct 2022 15:13
Operator : JC/MD
Sample : N5249-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-32-20221019

Quant Time: Oct 27 14:17:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

