

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032365.D
 Acq On : 25 Oct 2022 15:36
 Operator : JC/MD
 Sample : N5249-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-29-20221019

Quant Time: Oct 27 14:17:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127507	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	200798	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79856	48.143	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.280%		
35) Dibromofluoromethane	5.385	113	65250	46.698	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.400%		
50) Toluene-d8	8.653	98	234701	46.662	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.320%		
62) 4-Bromofluorobenzene	11.079	95	77418	42.033	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.060%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	4718	3.251	ug/l	89
32) 1,1,1-Trichloroethane	5.391	97	16104	7.672	ug/l	97
44) Trichloroethene	7.129	130	307946	214.663	ug/l	98
64) Tetrachloroethene	9.275	164	33766	29.821	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032365.D
Acq On : 25 Oct 2022 15:36
Operator : JC/MD
Sample : N5249-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-29-20221019

Quant Time: Oct 27 14:17:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

