

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032368.D
 Acq On : 25 Oct 2022 16:46
 Operator : JC/MD
 Sample : N5249-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-31-20221019

Quant Time: Oct 27 14:17:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126328	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197613	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	168372	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79743	48.523	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.040%		
35) Dibromofluoromethane	5.391	113	63363	46.078	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.160%		
50) Toluene-d8	8.653	98	232090	46.887	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	93.780%		
62) 4-Bromofluorobenzene	11.079	95	76843	42.393	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	84.780%		
Target Compounds						
27) cis-1,2-Dichloroethene	4.495	96	440	0.306	ug/l	80
44) Trichloroethene	7.129	130	54265	38.437	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032368.D
Acq On : 25 Oct 2022 16:46
Operator : JC/MD
Sample : N5249-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-31-20221019

Quant Time: Oct 27 14:17:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

