

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032371.D
 Acq On : 25 Oct 2022 17:56
 Operator : JC/MD
 Sample : N5249-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FSND-MW-24-20221019

Quant Time: Oct 27 14:18:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203104	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81280	49.613	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.220%
35) Dibromofluoromethane	5.391	113	65526	46.363	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.720%
50) Toluene-d8	8.653	98	236968	46.578	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.160%
62) 4-Bromofluorobenzene	11.079	95	81721	43.866	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.740%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	51577	45.885	ug/l	96
21) trans-1,2-Dichloroethene	3.093	96	2470	2.018	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	5916	4.127	ug/l	93
44) Trichloroethene	7.129	130	9103	6.273	ug/l	90
52) Toluene	8.726	92	2256	0.661	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032371.D
Acq On : 25 Oct 2022 17:56
Operator : JC/MD
Sample : N5249-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FSND-MW-24-20221019

Quant Time: Oct 27 14:18:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

