

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032383.D
 Acq On : 25 Oct 2022 22:59
 Operator : JC/MD
 Sample : VX1025WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1025WBL02

Quant Time: Oct 27 14:20:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205671	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181004	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82619	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81443	48.234	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.460%
35) Dibromofluoromethane	5.385	113	67006	46.818	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.640%
50) Toluene-d8	8.653	98	244361	47.432	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.860%
62) 4-Bromofluorobenzene	11.079	95	82487	43.724	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032383.D
Acq On : 25 Oct 2022 22:59
Operator : JC/MD
Sample : VX1025WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1025WBL02

Quant Time: Oct 27 14:20:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

