

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032386.D
 Acq On : 26 Oct 2022 00:09
 Operator : JC/MD
 Sample : PB148552TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148552TB

Quant Time: Oct 27 14:21:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121781	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195082	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	172016	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78746	49.706	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.420%
35) Dibromofluoromethane	5.391	113	62396	45.964	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.920%
50) Toluene-d8	8.653	98	232344	47.547	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.100%
62) 4-Bromofluorobenzene	11.079	95	79506	44.432	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.860%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11573	22.510	ug/l	98
20) Methylene Chloride	2.794	84	6706	5.080	ug/l	93
37) Ethyl Acetate	4.715	43	15507	9.070	ug/l #	88
43) Isopropyl Acetate	6.342	43	72018	26.837	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032386.D
Acq On : 26 Oct 2022 00:09
Operator : JC/MD
Sample : PB148552TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148552TB

Quant Time: Oct 27 14:21:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

