

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032389.D
 Acq On : 26 Oct 2022 01:19
 Operator : JC/MD
 Sample : N5267-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP2-1024

Quant Time: Oct 27 14:21:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	207089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183790	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84297	50.665	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.340%
35) Dibromofluoromethane	5.385	113	66749	46.319	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.640%
50) Toluene-d8	8.653	98	246920	47.600	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.200%
62) 4-Bromofluorobenzene	11.079	95	84311	44.385	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.760%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	14442	26.747	ug/l	99
20) Methylene Chloride	2.794	84	6647	4.794	ug/l	97
37) Ethyl Acetate	4.721	43	14464	7.969	ug/l	99
43) Isopropyl Acetate	6.342	43	66725	23.423	ug/l	100
95) Naphthalene	13.780	128	9857	1.692	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032389.D
Acq On : 26 Oct 2022 01:19
Operator : JC/MD
Sample : N5267-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP2-1024

Quant Time: Oct 27 14:21:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

