

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
 Data File : VX032391.D
 Acq On : 26 Oct 2022 02:06
 Operator : JC/MD
 Sample : N5267-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP4-1024

Quant Time: Oct 27 14:21:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128803	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206290	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186700	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87305	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83712	49.960	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.920%
35) Dibromofluoromethane	5.385	113	66193	46.111	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.220%
50) Toluene-d8	8.647	98	248630	48.115	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.240%
62) 4-Bromofluorobenzene	11.079	95	87028	45.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.980%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	22088	40.621	ug/l	98
20) Methylene Chloride	2.794	84	7528	5.392	ug/l	98
25) 2-Butanone	4.562	43	5340	6.290	ug/l	95
37) Ethyl Acetate	4.721	43	15426	8.532	ug/l	99
43) Isopropyl Acetate	6.342	43	69514	24.497	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032391.D
Acq On : 26 Oct 2022 02:06
Operator : JC/MD
Sample : N5267-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP4-1024

Quant Time: Oct 27 14:21:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

