

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102522\
 Data File : VX032353.D
 Acq On : 25 Oct 2022 10:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 14:15:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	108	0.00
2 T	Dichlorodifluoromethane	50.000	45.245	9.5	99	0.00
3 P	Chloromethane	50.000	41.796	16.4	85	0.00
4 C	Vinyl Chloride	50.000	43.689	12.6#	90	0.00
5 T	Bromomethane	50.000	44.783	10.4	93	0.00
6 T	Chloroethane	50.000	42.201	15.6	92	0.00
7 T	Trichlorofluoromethane	50.000	48.957	2.1	106	0.00
8 T	Diethyl Ether	50.000	47.763	4.5	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.940	0.1	110	0.00
10 T	Methyl Iodide	50.000	32.799	34.4#	66	0.00
11 T	Tert butyl alcohol	250.000	199.445	20.2	86	0.00
12 CM	1,1-Dichloroethene	50.000	46.606	6.8#	96	0.00
13 T	Acrolein	250.000	232.649	6.9	99	0.00
14 T	Allyl chloride	50.000	48.747	2.5	99	0.00
15 T	Acrylonitrile	250.000	224.842	10.1	92	0.00
16 T	Acetone	250.000	274.339	-9.7	117	0.00
17 T	Carbon Disulfide	50.000	44.313	11.4	90	0.00
18 T	Methyl Acetate	50.000	47.002	6.0	96	0.00
19 T	Methyl tert-butyl Ether	50.000	49.373	1.3	102	0.00
20 T	Methylene Chloride	50.000	43.834	12.3	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	45.816	8.4	93	0.00
22 T	Diisopropyl ether	50.000	48.904	2.2	101	0.00
23 T	Vinyl Acetate	250.000	242.830	2.9	98	0.00
24 P	1,1-Dichloroethane	50.000	48.735	2.5	102	0.00
25 T	2-Butanone	250.000	249.917	0.0	103	0.00
26 T	2,2-Dichloropropane	50.000	57.132	-14.3	113	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.410	3.2	98	0.00
28 T	Bromochloromethane	50.000	44.983	10.0	100	0.00
29 T	Tetrahydrofuran	250.000	226.782	9.3	92	0.00
30 C	Chloroform	50.000	49.409	1.2#	102	0.00
31 T	Cyclohexane	50.000	46.741	6.5	99	0.00
32 T	1,1,1-Trichloroethane	50.000	52.634	-5.3	106	0.00
33 S	1,2-Dichloroethane-d4	50.000	49.866	0.3	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	107	0.00
35 S	Dibromofluoromethane	50.000	49.394	1.2	100	0.00
36 T	1,1-Dichloropropene	50.000	48.250	3.5	99	0.00
37 T	Ethyl Acetate	50.000	45.778	8.4	95	0.00
38 T	Carbon Tetrachloride	50.000	53.044	-6.1	106	0.00
39 T	Methylcyclohexane	50.000	48.617	2.8	101	0.00
40 TM	Benzene	50.000	47.313	5.4	96	0.00
41 T	Methacrylonitrile	50.000	49.324	1.4	94	0.00
42 TM	1,2-Dichloroethane	50.000	51.899	-3.8	105	0.00
43 T	Isopropyl Acetate	50.000	50.135	-0.3	98	0.00
44 TM	Trichloroethene	50.000	48.013	4.0	97	0.00
45 C	1,2-Dichloropropane	50.000	49.921	0.2#	100	0.00
46 T	Dibromomethane	50.000	49.674	0.7	99	0.00
47 T	Bromodichloromethane	50.000	55.085	-10.2	109	0.00
48 T	Methyl methacrylate	50.000	50.685	-1.4	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102522\
 Data File : VX032353.D
 Acq On : 25 Oct 2022 10:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 14:15:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	864.194	13.6	84	0.00
50 S	Toluene-d8	50.000	48.161	3.7	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	243.073	2.8	95	0.00
52 CM	Toluene	50.000	48.588	2.8#	98	0.00
53 T	t-1,3-Dichloropropene	50.000	50.431	-0.9	108	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.888	-7.8	104	0.00
55 T	1,1,2-Trichloroethane	50.000	50.308	-0.6	100	0.00
56 T	Ethyl methacrylate	50.000	50.631	-1.3	100	0.00
57 T	1,3-Dichloropropane	50.000	50.373	-0.7	101	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	244.535	2.2	99	0.00
59 T	2-Hexanone	250.000	254.714	-1.9	100	0.00
60 T	Dibromochloromethane	50.000	56.357	-12.7	108	0.00
61 T	1,2-Dibromoethane	50.000	50.095	-0.2	99	0.00
62 S	4-Bromofluorobenzene	50.000	51.974	-3.9	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	106	0.00
64 T	Tetrachloroethene	50.000	47.679	4.6	98	0.00
65 PM	Chlorobenzene	50.000	48.947	2.1	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.528	-5.1	103	0.00
67 C	Ethyl Benzene	50.000	50.042	-0.1#	100	0.00
68 T	m/p-Xylenes	100.000	98.604	1.4	100	0.00
69 T	o-Xylene	50.000	50.428	-0.9	100	0.00
70 T	Styrene	50.000	50.366	-0.7	100	0.00
71 P	Bromoform	50.000	51.688	-3.4	111	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	110	0.00
73 T	Isopropylbenzene	50.000	48.162	3.7	103	0.00
74 T	N-ethyl acetate	50.000	49.407	1.2	100	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	46.331	7.3	99	0.00
76 T	1,2,3-Trichloropropane	50.000	48.108	3.8	103	0.00
77 T	Bromobenzene	50.000	47.092	5.8	101	0.00
78 T	n-propylbenzene	50.000	49.358	1.3	104	0.00
79 T	2-Chlorotoluene	50.000	48.242	3.5	103	0.00
80 T	1,3,5-Trimethylbenzene	50.000	48.654	2.7	104	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.213	1.6	108	0.00
82 T	4-Chlorotoluene	50.000	49.238	1.5	104	0.00
83 T	tert-Butylbenzene	50.000	49.126	1.7	105	0.00
84 T	1,2,4-Trimethylbenzene	50.000	49.923	0.2	104	0.00
85 T	sec-Butylbenzene	50.000	50.282	-0.6	106	0.00
86 T	p-Isopropyltoluene	50.000	50.511	-1.0	105	0.00
87 T	1,3-Dichlorobenzene	50.000	48.587	2.8	105	0.00
88 T	1,4-Dichlorobenzene	50.000	47.625	4.8	103	0.00
89 T	n-Butylbenzene	50.000	51.168	-2.3	106	0.00
90 T	Hexachloroethane	50.000	55.252	-10.5	114	0.00
91 T	1,2-Dichlorobenzene	50.000	48.894	2.2	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.445	-4.9	103	0.00
93 T	1,2,4-Trichlorobenzene	50.000	47.988	4.0	102	0.00
94 T	Hexachlorobutadiene	50.000	47.716	4.6	104	0.00
95 T	Naphthalene	50.000	47.896	4.2	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032353.D
Acq On : 25 Oct 2022 10:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 27 14:15:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.378	5.2	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6