

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102522\
 Data File : VX032395.D
 Acq On : 26 Oct 2022 03:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 14:22:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	45.838	8.3	94	0.00
3 P	Chloromethane	50.000	43.651	12.7	83	0.00
4 C	Vinyl Chloride	50.000	45.779	8.4#	88	0.00
5 T	Bromomethane	50.000	51.004	-2.0	99	0.00
6 T	Chloroethane	50.000	35.790	28.4#	74	0.00
7 T	Trichlorofluoromethane	50.000	49.863	0.3	101	0.00
8 T	Diethyl Ether	50.000	50.764	-1.5	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.933	2.1	101	0.00
10 T	Methyl Iodide	50.000	37.234	25.5#	70	0.00
11 T	Tert butyl alcohol	250.000	272.684	-9.1	107	0.00
12 CM	1,1-Dichloroethene	50.000	48.226	3.5#	93	0.00
13 T	Acrolein	250.000	173.535	30.6#	70	0.00
14 T	Allyl chloride	50.000	50.344	-0.7	96	0.00
15 T	Acrylonitrile	250.000	257.198	-2.9	99	0.00
16 T	Acetone	250.000	266.649	-6.7	107	0.00
17 T	Carbon Disulfide	50.000	42.008	16.0	80	0.00
18 T	Methyl Acetate	50.000	55.895	-11.8	107	0.00
19 T	Methyl tert-butyl Ether	50.000	53.902	-7.8	104	0.00
20 T	Methylene Chloride	50.000	47.134	5.7	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.865	4.3	91	0.00
22 T	Diisopropyl ether	50.000	52.681	-5.4	102	0.00
23 T	Vinyl Acetate	250.000	262.918	-5.2	99	0.00
24 P	1,1-Dichloroethane	50.000	51.958	-3.9	102	0.00
25 T	2-Butanone	250.000	267.062	-6.8	103	0.00
26 T	2,2-Dichloropropane	50.000	42.612	14.8	79	0.00
27 T	cis-1,2-Dichloroethene	50.000	51.771	-3.5	98	0.00
28 T	Bromochloromethane	50.000	50.255	-0.5	104	0.00
29 T	Tetrahydrofuran	250.000	264.311	-5.7	101	0.00
30 C	Chloroform	50.000	52.813	-5.6#	103	0.00
31 T	Cyclohexane	50.000	48.027	3.9	95	0.00
32 T	1,1,1-Trichloroethane	50.000	55.891	-11.8	105	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.842	-7.7	104	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	105	0.00
35 S	Dibromofluoromethane	50.000	49.666	0.7	99	0.00
36 T	1,1-Dichloropropene	50.000	47.902	4.2	97	0.00
37 T	Ethyl Acetate	50.000	49.438	1.1	101	0.00
38 T	Carbon Tetrachloride	50.000	52.048	-4.1	102	0.00
39 T	Methylcyclohexane	50.000	45.481	9.0	93	0.00
40 TM	Benzene	50.000	47.865	4.3	96	0.00
41 T	Methacrylonitrile	50.000	53.442	-6.9	100	0.00
42 TM	1,2-Dichloroethane	50.000	52.499	-5.0	104	0.00
43 T	Isopropyl Acetate	50.000	53.415	-6.8	103	0.00
44 TM	Trichloroethene	50.000	48.656	2.7	97	0.00
45 C	1,2-Dichloropropane	50.000	50.658	-1.3#	100	0.00
46 T	Dibromomethane	50.000	50.922	-1.8	100	0.00
47 T	Bromodichloromethane	50.000	54.070	-8.1	105	0.00
48 T	Methyl methacrylate	50.000	55.129	-10.3	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX102522\
 Data File : VX032395.D
 Acq On : 26 Oct 2022 03:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 27 14:22:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 22 04:56:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1038.636	-3.9	100	0.00
50 S	Toluene-d8	50.000	48.686	2.6	95	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.102	-7.2	103	0.00
52 CM	Toluene	50.000	48.747	2.5#	97	0.00
53 T	t-1,3-Dichloropropene	50.000	47.809	4.4	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	51.374	-2.7	98	0.00
55 T	1,1,2-Trichloroethane	50.000	52.071	-4.1	102	0.00
56 T	Ethyl methacrylate	50.000	54.164	-8.3	105	0.00
57 T	1,3-Dichloropropane	50.000	51.661	-3.3	102	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	235.815	5.7	94	0.00
59 T	2-Hexanone	250.000	268.096	-7.2	103	0.00
60 T	Dibromochloromethane	50.000	56.254	-12.5	106	0.00
61 T	1,2-Dibromoethane	50.000	52.144	-4.3	102	0.00
62 S	4-Bromofluorobenzene	50.000	53.037	-6.1	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	105	0.00
64 T	Tetrachloroethene	50.000	47.311	5.4	97	0.00
65 PM	Chlorobenzene	50.000	49.404	1.2	100	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.115	-6.2	103	0.00
67 C	Ethyl Benzene	50.000	49.841	0.3#	99	0.00
68 T	m/p-Xylenes	100.000	97.638	2.4	98	0.00
69 T	o-Xylene	50.000	50.469	-0.9	100	0.00
70 T	Styrene	50.000	50.709	-1.4	101	0.00
71 P	Bromoform	50.000	50.398	-0.8	108	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	107	0.00
73 T	Isopropylbenzene	50.000	48.950	2.1	102	0.00
74 T	N-ethyl acetate	50.000	53.826	-7.7	105	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	49.499	1.0	103	0.00
76 T	1,2,3-Trichloropropane	50.000	50.925	-1.8	105	0.00
77 T	Bromobenzene	50.000	49.510	1.0	103	0.00
78 T	n-propylbenzene	50.000	49.404	1.2	101	0.00
79 T	2-Chlorotoluene	50.000	50.072	-0.1	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	49.655	0.7	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.457	7.1	99	0.00
82 T	4-Chlorotoluene	50.000	50.420	-0.8	103	0.00
83 T	tert-Butylbenzene	50.000	50.044	-0.1	104	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.678	-1.4	102	0.00
85 T	sec-Butylbenzene	50.000	50.005	-0.0	102	0.00
86 T	p-Isopropyltoluene	50.000	50.637	-1.3	103	0.00
87 T	1,3-Dichlorobenzene	50.000	48.613	2.8	102	0.00
88 T	1,4-Dichlorobenzene	50.000	48.376	3.2	102	0.00
89 T	n-Butylbenzene	50.000	48.865	2.3	98	0.00
90 T	Hexachloroethane	50.000	53.601	-7.2	108	0.00
91 T	1,2-Dichlorobenzene	50.000	49.613	0.8	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	57.248	-14.5	109	0.00
93 T	1,2,4-Trichlorobenzene	50.000	46.813	6.4	96	0.00
94 T	Hexachlorobutadiene	50.000	44.390	11.2	94	0.00
95 T	Naphthalene	50.000	50.349	-0.7	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032395.D
Acq On : 26 Oct 2022 03:39
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 27 14:22:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 22 04:56:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.185	5.6	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6