



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102522\
Data File : VX032380.D
Acq On : 25 Oct 2022 21:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101922W.M
Title : SW846 8260
Last Update : Sat Oct 22 04:56:40 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.4	6539	PASS
75	95	30	60	54.5	16638	PASS
95	95	100	100	100.0	30539	PASS
96	95	5	9	6.6	2020	PASS
173	174	0.00	2	0.9	217	PASS
174	95	50	100	77.4	23642	PASS
175	174	5	9	7.7	1828	PASS
176	174	95	101	95.9	22675	PASS
177	176	5	9	7.0	1588	PASS