

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038471.D
 Acq On : 25 Oct 2023 11:52
 Operator : JC/MD
 Sample : 05007-20DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-26-23-26DL

Quant Time: Oct 26 04:02:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	177178	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	304316	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303555	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	110548	38.792	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	77.580%#
35) Dibromofluoromethane	5.385	113	98964	45.132	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.260%
50) Toluene-d8	8.647	98	370753	44.247	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.500%#
62) 4-Bromofluorobenzene	11.079	95	142926	44.993	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.980%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	2732	1.235	ug/l	92
64) Tetrachloroethene	9.275	164	11345	5.773	ug/l	92
93) 1,2,4-Trichlorobenzene	13.585	180	112255	32.672	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	28190	8.474	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038471.D
Acq On : 25 Oct 2023 11:52
Operator : JC/MD
Sample : 05007-20DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-26-23-26DL

Quant Time: Oct 26 04:02:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

