

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038480.D
 Acq On : 25 Oct 2023 15:20
 Operator : JC/MD
 Sample : 04980-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BRIDAL-SHOP-BAY-SUMP

Quant Time: Oct 26 04:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	161847	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	280044	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	285681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	155330	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103270	39.670	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.340%
35) Dibromofluoromethane	5.391	113	92270	45.726	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.460%
50) Toluene-d8	8.647	98	344276	44.648	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.300%#
62) 4-Bromofluorobenzene	11.079	95	133988	45.835	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.660%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.141	130	923	0.453	ug/l	96
64) Tetrachloroethene	9.275	164	170913	92.415	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038480.D
Acq On : 25 Oct 2023 15:20
Operator : JC/MD
Sample : 04980-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BRIDAL-SHOP-BAY-SUMP

Quant Time: Oct 26 04:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

