

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038481.D
 Acq On : 25 Oct 2023 15:43
 Operator : JC/MD
 Sample : 04980-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MAIN-BAILER-SUMP

Quant Time: Oct 26 04:06:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163106	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	282557	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286317	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104092	39.677	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.360%
35) Dibromofluoromethane	5.385	113	92873	45.615	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.240%
50) Toluene-d8	8.647	98	340536	43.770	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.540%#
62) 4-Bromofluorobenzene	11.079	95	133805	45.365	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.740%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.489	96	15665	7.119	ug/l	78
44) Trichloroethene	7.129	130	3395	1.653	ug/l	89
64) Tetrachloroethene	9.275	164	24685	13.318	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038481.D
Acq On : 25 Oct 2023 15:43
Operator : JC/MD
Sample : 04980-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MAIN-BAILER-SUMP

Quant Time: Oct 26 04:06:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

