

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038482.D
 Acq On : 25 Oct 2023 16:06
 Operator : JC/MD
 Sample : 05007-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-24-10-13

Quant Time: Oct 26 04:06:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	164935	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	288960	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	297136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	159735	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	107839	40.650	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.300%
35) Dibromofluoromethane	5.385	113	94864	45.561	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.120%
50) Toluene-d8	8.647	98	352975	44.364	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.720%#
62) 4-Bromofluorobenzene	11.079	95	137860	45.704	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.400%
Target Compounds						
52) Toluene	8.720	92	12612	2.411	ug/l	92
64) Tetrachloroethene	9.275	164	2005	1.042	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038482.D
Acq On : 25 Oct 2023 16:06
Operator : JC/MD
Sample : 05007-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-24-10-13

Quant Time: Oct 26 04:06:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

