

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038484.D
 Acq On : 25 Oct 2023 16:53
 Operator : JC/MD
 Sample : 05038-05 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1067

Quant Time: Oct 26 04:07:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	164531	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	282975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	291933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	162114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	104954	39.660	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.320%
35) Dibromofluoromethane	5.385	113	93171	45.694	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.380%
50) Toluene-d8	8.647	98	346889	44.521	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.040%#
62) 4-Bromofluorobenzene	11.079	95	136431	46.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.380%
Target Compounds						
						Qvalue
25) 2-Butanone	4.568	43	5414	3.038	ug/l	94
51) 4-Methyl-2-Pentanone	8.574	43	1768	0.514	ug/l	98
59) 2-Hexanone	9.445	43	1375	0.497	ug/l	83
64) Tetrachloroethene	9.275	164	685	0.362	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038484.D
Acq On : 25 Oct 2023 16:53
Operator : JC/MD
Sample : 05038-05 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1067

Quant Time: Oct 26 04:07:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

