

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038485.D
 Acq On : 25 Oct 2023 17:16
 Operator : JC/MD
 Sample : 05038-07 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 AUD-1069

Quant Time: Oct 26 04:07:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	161376	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	278423	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	286326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	156695	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	102066	39.322	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	78.640%
35) Dibromofluoromethane	5.391	113	92770	46.241	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.480%
50) Toluene-d8	8.647	98	341706	44.573	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.140%#
62) 4-Bromofluorobenzene	11.079	95	134862	46.403	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.800%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8761	8.087	ug/l	90
20) Methylene Chloride	2.794	84	1378	0.676	ug/l	# 68
37) Ethyl Acetate	4.727	43	5927	1.804	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	12028	1.331	ug/l	# 35
85) sec-Butylbenzene	11.756	105	12028	1.048	ug/l	# 57

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038485.D
Acq On : 25 Oct 2023 17:16
Operator : JC/MD
Sample : 05038-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
AUD-1069

Quant Time: Oct 26 04:07:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

