

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038505.D
 Acq On : 26 Oct 2023 07:53
 Operator : JC/MD
 Sample : 05049-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B4

Quant Time: Oct 26 08:40:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	184533	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	318559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326175	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113671	38.298	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	76.600%#
35) Dibromofluoromethane	5.385	113	103024	44.882	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.760%
50) Toluene-d8	8.646	98	384838	43.874	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	87.740%#
62) 4-Bromofluorobenzene	11.079	95	151935	45.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.380%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	7317	5.907	ug/l	# 86
52) Toluene	8.720	92	2355	0.408	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038505.D
Acq On : 26 Oct 2023 07:53
Operator : JC/MD
Sample : 05049-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B4

Quant Time: Oct 26 08:40:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

