

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038509.D
 Acq On : 26 Oct 2023 09:26
 Operator : JC/MD
 Sample : 05065-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-01S

Quant Time: Oct 26 13:38:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	178502	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	308187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315404	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113117	39.399	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	78.800%
35) Dibromofluoromethane	5.385	113	103672	46.685	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.360%
50) Toluene-d8	8.647	98	377575	44.495	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.000%#
62) 4-Bromofluorobenzene	11.079	95	147341	45.800	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.600%
Target Compounds						
					Qvalue	
16) Acetone	2.374	43	2759	2.303	ug/l	99
20) Methylene Chloride	2.794	84	1239	0.550	ug/l	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038509.D
Acq On : 26 Oct 2023 09:26
Operator : JC/MD
Sample : 05065-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-01S

Quant Time: Oct 26 13:38:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

