

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
 Data File : VX038510.D
 Acq On : 26 Oct 2023 09:49
 Operator : JC/MD
 Sample : 05065-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-01i

Quant Time: Oct 26 13:38:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	174505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	301459	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	308989	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	170470	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	111264	39.641	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	79.280%
35) Dibromofluoromethane	5.391	113	99205	45.670	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.340%
50) Toluene-d8	8.647	98	366281	44.127	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.260%#
62) 4-Bromofluorobenzene	11.079	95	143666	45.654	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.300%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2298	1.962	ug/l #	87
20) Methylene Chloride	2.788	84	1182	0.536	ug/l	96
44) Trichloroethene	7.129	130	10527	4.804	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102523\
Data File : VX038510.D
Acq On : 26 Oct 2023 09:49
Operator : JC/MD
Sample : 05065-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-01i

Quant Time: Oct 26 13:38:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

